P1689



Thermo Fisher SCIENTIFIC

The world leader in serving science

HEIDEL BERG

MVZ Labor Dr. Limbach

Performance of a Multiplex Real-Time PCR Kit for Detection and Differentiation of Enteric Bacterial Pathogens

Jasmin Köffer¹, Sean Gregory², Bahareh Mohajer², Brittney Lehmann², Xuemei Wang², Camilla Ulekleiv², Oceane Sorel², Jelena Feenstra², Manoj Gandhi², Ulrich Eigner¹

¹MVZ Labor Dr. Limbach & Kollegen GbR, Heidelberg, Germany; ²Thermo Fisher Scientific, South San Francisco, CA, USA

Background

Bacterial infections are among the most common causes of infective enteritis worldwide. Stool culture is still commonly performed but it can be cumbersome and may not be sensitive enough to identify the causative pathogen. TaqPath[™] Enteric Bacterial Select Panel is a CE-IVD multiplex real-time PCR kit designed for diagnostic testing of most common enteric bacterial pathogens in one reaction: Campylobacter (jejuni, coli, upsaliensis); Salmonella spp. and Shigella spp./Enteroinvasive E. Coli (EIEC). This study evaluated the performance of the TaqPath™ **Enteric Bacterial Select Panel.**

Methods & Results

Preclinical performance evaluation

Table 1. Test performance on contrived samples

Contrived samples were created by inoculating cultures of C. jejuni, S. typhimurium and S. sonnei individually into negative stool samples at 2 concentrations (~3-5x limit of detection (LOD) and 10xLOD of the TaqPath™ Enteric Bacterial Select Panel) in 15 replicates. Contrived positive samples and 30 negative samples were tested in parallel using TaqPath[™] Enteric Bacterial Select Panel, RIDA[®]GENE Bacterial Stool Panel I and Allplex[™] GI-EB Screening kits. The results of the testing are shown in Table 1.

Target	TaqPath Enteric Bacterial Select Panel	RIDA GENE Bacterial Stool Panel I	Seegene Allplex GI- EB kit		
Campylobacter	100% (30/30)	100% (30/30)	43% (13/30)		
Salmonella	97% (29/30)	100% (30/30)	10% (3/30)		
Shigella	100% (30/30)	100% (30/30)	100% (30/30)		
Negative	100% (30/30)	100% (30/30)	100% (30/30)		

Contrived sample analysis demonstrated higher analytical sensitivity of the TaqPath[™] Enteric Bacterial Select Panel and RIDA[®]GENE Bacterial Stool Panel I for C. jejuni and S. typhimurium compared to Allplex[™] GI-EB Screening kit, while all 3 tests showed comparable performance for detection of S. sonnei.

Clinical performance evaluation I

In total 407 stool specimens from symptomatic individuals collected in Germany, France and Ivory Coast were tested using TaqPath[™] Enteric Bacterial Select Panel and RIDA[®] GENE Bacterial Stool Panel I and discordant sample resolution was performed using different sets of primers and probes for each target (Figure 1.). Samples with invalid results were excluded (N=12). Study results are summarized in Table 2. Co-infections were detected in 21 samples.



Table 2. Clinical performance of the TaqPath[™] Enteric Bacterial Select Panel

	Consensus result									
	Campylobacter spp.		Salmonella spp.			Shigella spp./ EIEC				
		Positive	Negative	Total	Positive	Negative	Total	Positive	Negative	Total
TaqPath [™] Enteric Bacterial	Positive	93	0	93	77	7	84	43	0	43
	Negative	2	300	302	0	311	311	0	352	352
Select Panel	Total	95	300	395	77	318	395	43	352	395
Clinical sensitivity (95% CI)		98% (92.6% - 99.7%)			100% (95.3% - 100.0%)			100% (91.8% - 100.0%)		
Clinical specificity (95% CI)		100%	6 (98.8% - 10	0.0%)	98% (95.5% - 99.1%) 100% (99.0% - 10			0.0%)		

Clinical sensitivity of the TaqPath[™] Enteric Bacterial Select Panel was 100% for *Salmonella spp*. and Shigella spp./EIEC, and 98% for Campylobacter; while specificity was 98%, 100% and 100%, respectively.

Clinical performance evaluation II

Testing of 217 stool samples from symptomatic individuals was performed in parallel using the TaqPath[™] Enteric Bacterial Select Panel and BD MAX[™] Enteric Bacterial Panel (Figure 2.). Two samples were excluded due to invalid results. The result summary is shown in Table 3.





Table 3. Performance comparison of TaqPath[™] Enteric Bacterial Select Panel and BD MAX[™] Enteric Bacterial Panel

	BD MAX [™] Enteric Bacterial Panel									
		Campylobacter spp.			Salmonella spp.			Shigella spp./ EIEC		
		Positive	Negative	Total	Positive	Negative	Total	Positive	Negative	Total
TaqPath™ Enteric Bacterial Select Panel	Positive	56	5	61	57	1	58	31	1	32
	Negative	0	154	154	1	156	157	0	183	183
	Total	56	159	215	58	157	215	31	184	215
Positive percent agreem	100% (93.6% - 100.0%) 98% (90.9% - 99.7%) 100% (89			39.0% - 100.0%)						
Negative percent agreen	97% (92.9% - 9	98.6%)	99% (96.5% - 99.9%) 99% (97.0% - 99.9%			9.9%)			

For 3/6 discordant samples routine culture result was available and showed *Campylobacter* positive result in agreement with the TaqPath[™] Enteric Bacterial Select Panel. Both molecular tests detected *Campylobacter* in one and *Salmonella* in another sample characterized as

Clinical performance evaluation against routine stool culture

In total 500 stool samples collected from January to March 2023 in Germany from symptomatic patients were tested with TaqPath[™] Enteric Bacterial Select Panel and routine stool culture for Salmonella, Shigella and Campylobacter. MALDI-TOF MS (Bruker) was used to identify colonies suspected of Salmonella and Campylobacter. All samples were also tested with the RIDASCREEN[®] Campylobacter antigen test (R-Biopharm). Shigella spp. was identified by biochemical (api20, api50 CHE, bioMérieux) and serological tests. Discordant samples were tested with 2 resolver methods (Figure 3.).

Figure 3. Study design of the clinical performance evaluation against routine stool culture



Salmonella was detected in 2/500 samples using both routine culture and the TaqPath[™] Enteric Panel. Campylobacter was detected in 26/500 samples with the TaqPath[™] kit, while routine culture detected only 46.1% (12/26) of these infections. The results obtained by the TaqPath[™] kit were shown to be true positive for *Campylobacter* using either BD MAXTM Bacterial Panel or BIOFIRE® Enteric FILMARRAY[®] GI Panel (Table 4.). Four samples showing negative result with BD MAX and culture, but positive with BIOFIRE and TaqPath

ample	TaqPath™ Enteric Bacterial Select Panel		Routine culture	RIDASCREEN®	BD MAX [™] Enteric Bacterial	BIOFIRE® FILMARRAY®	
U	Camp. Result	Camp. Ct value	result	antigen test	Panel	GI Panel	
1	Negative	-	Positive	Negative	Negative	N/A 🗸	
2	Positive	19.36	Negative	Negative	Negative	Positive 🗸	
3	Positive	26.64	Negative	Negative	Negative	Positive 🗸	
4	Positive	28.88	Negative	Negative	Negative	Positive 🗸	
5	Positive	35.81	Negative	Negative	Negative	Positive 🗸	
6	Positive	25.47	Negative	Negative	Positive	N/A 🗸	
7	Positive	26.19	Negative	Negative	Positive	N/A 🗸	
8	Positive	26.91	Negative	Negative	Positive	N/A 🗸	
9	Positive	28.54	Negative	Negative	Positive	N/A 🗸	
10	Positive	31.71	Negative	Negative	Positive	N/A 🗸	
11	Positive	25.88	Negative	Positive	Positive	N/A 🗸	
12	Positive	22.58	Negative	Positive	Positive	N/A 🗸	
13	Positive	29.49	Negative	Positive	Positive	N/A 🗸	
14	Positive	19.05	Negative	Positive	Positive	N/A 🗸	

negative using routine culture.	

are likely to represent *C. upsaliensis*.

N/A 🚽 23.11 15 Positive Negative Positive Positive

N/A, results not available as samples were not tested with this method

Enteric Bacterial Select Panel shows comparable TagPath[™] performance to the BD MAXTM Enteric Bacterial Panel with high positive and negative percent agreement for all 3 targets.

TaqPath[™] Enteric Bacterial Select Panel shows significantly higher clinical sensitivity compared to routine culture for detection of *Campylobacter*, as routine culture testing did not identify >50% of infections.

Conclusion

The TaqPath[™] Enteric Bacterial Select Panel is a highly accurate method for quick (<2 hours eluate to result turnaround time) detection and differentiation of most common enteric bacterial pathogens. Compared to culture-based approaches, molecular testing offers higher sensitivity and reduced time-to-result for diagnostic testing of bacterial enteric infections.

CE-IVD In Vitro Diagnostic Use

© 2023 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified. Ridagene and RIDASCREEN are trademarks of R-Biopharm AG. BioFire and FILMARRAY are trademarks of BioFire Diagnostics, Inc. ALLPLEX is a trademark of Seegene Inc. BD MAX is a trademark of Becton, Dickinson and Company. API is a trademark of bioMérieux.